

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARZL-0146  
Location: WALL TOWNSHIP POLICE  
Calibration File No.: 01254 Calib. Date: 01/25/2017 Calib. No.: 00026  
Certification File No.: 01255 Cert. Date: 01/25/2017 Cert. No.: 00022  
Linearity File No.: 01256 Lin. Date: 01/25/2017 Lin. No.: 00022  
Solution File No.: 01323 Soln. Date: 06/30/2017 Soln. No.: 00244  
Sequential File No.: 01323 File Date: 06/30/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-114  
Control Solution %: 0.100% Expires: 10/28/2018  
Solution Control Lot: 16330 Bottle No.: 0296

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 14:24D        |                               |                           |
| Control 1 EC      | 0.101%         | 14:25D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.101%         | 14:25D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 14:26D        |                               |                           |
| Control 2 EC      | 0.100%         | 14:26D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.101%         | 14:26D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 14:27D        |                               |                           |
| Control 3 EC      | 0.100%         | 14:28D        | 34.0°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.100%         | 14:28D        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 14:28D        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: DK #129

Badge No.: 129

Date: 06/30/2017