

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01337 Calib. Date: 07/21/2017 Calib. No.: 00027
Certification File No.: 01338 Cert. Date: 07/21/2017 Cert. No.: 00023
Linearity File No.: 01339 Lin. Date: 07/21/2017 Lin. No.: 00023
Solution File No.: 01346 Soln. Date: 08/04/2017 Soln. No.: 00247
Sequential File No.: 01346 File Date: 08/04/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0015
Control Solution %: 0.100% Expires: 01/30/2019
Solution Control Lot: 17050 Bottle No.: 1232

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:02D		
Control 1 EC	0.099%	09:03D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	09:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:04D		
Control 2 EC	0.098%	09:04D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	09:04D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:05D		
Control 3 EC	0.098%	09:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:06D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A.

Signature: 

Badge No.: 129

Date: 08/04/2017