

# Calibrating Unit

## New Standard Solution Report

|                         |                       |              |            |
|-------------------------|-----------------------|--------------|------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C | Serial No.:  | ARXA-0037  |
| Location:               | WALL TOWNSHIP POLICE  |              |            |
| Calibration File No.:   | 01678                 | Calib. Date: | 07/21/2017 |
| Certification File No.: | 01679                 | Cert. Date:  | 07/21/2017 |
| Linearity File No.:     | 01680                 | Lin. Date:   | 07/21/2017 |
| Solution File No.:      | 01688                 | Soln. Date:  | 08/07/2017 |
| Sequential File No.:    | 01688                 | File Date:   | 08/07/2017 |
|                         |                       | Calib. No.:  | 00033      |
|                         |                       | Cert. No.:   | 00024      |
|                         |                       | Lin. No.:    | 00024      |
|                         |                       | Soln. No.:   | 00291      |
| Calibrating Unit:       | WET                   | Model No.:   | CU-34      |
| Control Solution %:     | 0.100%                | Serial No.:  | DDAB-0014  |
| Solution Control Lot:   | 17050                 | Expires:     | 01/30/2019 |
|                         |                       | Bottle No.:  | 1219       |

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 14:30D        |                               |                           |
| Control 1 EC      | 0.099%         | 14:31D        | 33.9°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.098%         | 14:31D        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 14:31D        |                               |                           |
| Control 2 EC      | 0.097%         | 14:32D        | 33.9°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.099%         | 14:32D        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 14:33D        |                               |                           |
| Control 3 EC      | 0.097%         | 14:33D        | 33.9°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.100%         | 14:33D        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 14:34D        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: PTL. #129

Badge No.: 129

Date: 08/07/2017