

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01337	Calib. Date:	07/21/2017
Certification File No.:	01338	Cert. Date:	07/21/2017
Linearity File No.:	01339	Lin. Date:	07/21/2017
Solution File No.:	01351	Soln. Date:	08/18/2017
Sequential File No.:	01351	File Date:	08/18/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	17050	Expires:	01/30/2019
		Bottle No.:	1314

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:53D		
Control 1 EC	0.099%	08:53D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:54D		
Control 2 EC	0.099%	08:55D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:55D		
Control 3 EC	0.099%	08:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: 

Badge No.: 129

Date: 08/18/2017