

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01678
Certification File No.: 01679
Linearity File No.: 01680
Solution File No.: 01703
Sequential File No.: 01703

Serial No.: ARXA-0037

Calib. Date: 07/21/2017
Cert. Date: 07/21/2017
Lin. Date: 07/21/2017
Soln. Date: 09/22/2017
File Date: 09/22/2017

Calib. No.: 00033
Cert. No.: 00024
Lin. No.: 00024
Soln. No.: 00296

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 17050

Model No.: CU-34

Serial No.: DDAB-0014
Expires: 01/30/2019
Bottle No.: 1211

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:14D		
Control 1 EC	0.099%	14:14D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:14D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:15D		
Control 2 EC	0.097%	14:16D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	14:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:16D		
Control 3 EC	0.097%	14:17D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:18D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: 

Badge No.: 129

Date: 09/22/2017