

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01337 **Calib. Date:** 07/21/2017 **Calib. No.:** 00027
Certification File No.: 01338 **Cert. Date:** 07/21/2017 **Cert. No.:** 00023
Linearity File No.: 01339 **Lin. Date:** 07/21/2017 **Lin. No.:** 00023
Solution File No.: 01375 **Soln. Date:** 09/30/2017 **Soln. No.:** 00252
Sequential File No.: 01375 **File Date:** 09/30/2017

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0015
Control Solution %: 0.100% **Expires:** 01/30/2019
Solution Control Lot: 17050 **Bottle No.:** 1237

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:59D		
Control 1 EC	0.098%	11:00D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:00D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:00D		
Control 2 EC	0.098%	11:01D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:02D		
Control 3 EC	0.098%	11:02D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: 

Badge No.: 129

Date: 09/30/2017