

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01678	Calib. Date:	07/21/2017
Certification File No.:	01679	Cert. Date:	07/21/2017
Linearity File No.:	01680	Lin. Date:	07/21/2017
Solution File No.:	01726	Soln. Date:	11/12/2017
Sequential File No.:	01726	File Date:	11/12/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	17290	Expires:	09/14/2019
		Bottle No.:	1091

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:06S		
Control 1 EC	0.097%	11:06S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	11:06S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:07S		
Control 2 EC	0.095%	11:08S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	11:08S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:09S		
Control 3 EC	0.096%	11:09S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	11:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:10S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DRACKWICZ

First Name: AMY

MI: M

Signature: Agt [Signature] 177

Badge No.: 177

Date: 11/12/2017