

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0037  
**Location:** WALL TOWNSHIP POLICE  
**Calibration File No.:** 01678 **Calib. Date:** 07/21/2017 **Calib. No.:** 00033  
**Certification File No.:** 01679 **Cert. Date:** 07/21/2017 **Cert. No.:** 00024  
**Linearity File No.:** 01680 **Lin. Date:** 07/21/2017 **Lin. No.:** 00024  
**Solution File No.:** 01728 **Soln. Date:** 11/25/2017 **Soln. No.:** 00303  
**Sequential File No.:** 01728 **File Date:** 11/25/2017

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDAB-0014  
**Control Solution %:** 0.100% **Expires:** 09/14/2019  
**Solution Control Lot:** 17290 **Bottle No.:** 1071

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 07:23S        |                               |                           |
| Control 1 EC      | 0.099%         | 07:24S        | 33.9°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.098%         | 07:24S        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 07:25S        |                               |                           |
| Control 2 EC      | 0.097%         | 07:25S        | 33.9°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.097%         | 07:25S        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 07:26S        |                               |                           |
| Control 3 EC      | 0.096%         | 07:27S        | 33.9°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.098%         | 07:27S        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 07:27S        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


### Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature:



Badge No.: 129

Date: 11/25/2017