

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01678 Calib. Date: 07/21/2017 Calib. No.: 00033
Certification File No.: 01679 Cert. Date: 07/21/2017 Cert. No.: 00024
Linearity File No.: 01680 Lin. Date: 07/21/2017 Lin. No.: 00024
Solution File No.: 01739 Soln. Date: 12/11/2017 Soln. No.: 00307
Sequential File No.: 01739 File Date: 12/11/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0014
Control Solution %: 0.100% Expires: 09/14/2019
Solution Control Lot: 17290 Bottle No.: 1174

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:58S		
Control 1 EC	0.097%	04:58S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	04:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:59S		
Control 2 EC	0.096%	04:59S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	04:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:00S		
Control 3 EC	0.095%	05:01S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	05:01S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:02S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MARINO

First Name: ANTHONY

MI: T

Signature:  #181

Badge No.: 181

Date: 12/11/2017