

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01678 **Calib. Date:** 07/21/2017 **Calib. No.:** 00033
Certification File No.: 01679 **Cert. Date:** 07/21/2017 **Cert. No.:** 00024
Linearity File No.: 01680 **Lin. Date:** 07/21/2017 **Lin. No.:** 00024
Solution File No.: 01745 **Soln. Date:** 12/25/2017 **Soln. No.:** 00309
Sequential File No.: 01745 **File Date:** 12/25/2017

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0014
Control Solution %: 0.100% **Expires:** 09/14/2019
Solution Control Lot: 17290 **Bottle No.:** 0323

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:23S		
Control 1 EC	0.097%	00:23S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	00:23S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:24S		
Control 2 EC	0.096%	00:25S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	00:25S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:25S		
Control 3 EC	0.096%	00:26S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:27S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MAZZACCARO

First Name: VINCENT

MI: J

Signature: Sgt. Mazza #122

Badge No.: 172

Date: 12/25/2017

EMCU' 12/26