

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01746 **Calib. Date:** 12/26/2017 **Calib. No.:** 00034
Certification File No.: 01747 **Cert. Date:** 12/26/2017 **Cert. No.:** 00025
Linearity File No.: 01748 **Lin. Date:** 12/26/2017 **Lin. No.:** 00025
Solution File No.: 01758 **Soln. Date:** 01/25/2018 **Soln. No.:** 00312
Sequential File No.: 01758 **File Date:** 01/25/2018

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXB S3-113
Control Solution %: 0.100% **Expires:** 09/14/2019
Solution Control Lot: 17290 **Bottle No.:** 0211

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:17S		
Control 1 EC	0.100%	13:18S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	13:18S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:19S		
Control 2 EC	0.099%	13:19S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:19S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:20S		
Control 3 EC	0.098%	13:21S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:21S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:21S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: 

Badge No.: 129

Date: 01/25/2018