

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01746	Calib. Date:	12/26/2017
Certification File No.:	01747	Cert. No.:	00034
Linearity File No.:	01748	Cert. Date:	12/26/2017
Solution File No.:	01761	Lin. Date:	12/26/2017
Sequential File No.:	01761	Lin. No.:	00025
		Soln. Date:	02/08/2018
		Soln. No.:	00313
		File Date:	02/08/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-113
Solution Control Lot:	17290	Expires:	09/14/2019
		Bottle No.:	1134

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:34S		
Control 1 EC	0.103%	09:35S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:35S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:36S		
Control 2 EC	0.100%	09:36S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:36S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:37S		
Control 3 EC	0.099%	09:38S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:38S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:38S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: PTL. K #129

Badge No.: 129

Date: 02/08/2018