

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01404	Calib. Date:	12/26/2017
Certification File No.:	01405	Cert. Date:	12/26/2017
Linearity File No.:	01406	Lin. Date:	12/26/2017
Solution File No.:	01415	Soln. Date:	02/08/2018
Sequential File No.:	01415	File Date:	02/08/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-114
Solution Control Lot:	17290	Expires:	09/14/2019
		Bottle No.:	0209

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:38S		
Control 1 EC	0.098%	09:39S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:39S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:40S		
Control 2 EC	0.098%	09:40S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:40S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:41S		
Control 3 EC	0.099%	09:42S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:42S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:42S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: _____

PLK #129

Badge No.: 129

Date: 02/08/2018