

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01746 Calib. Date: 12/26/2017 Calib. No.: 00034
Certification File No.: 01747 Cert. Date: 12/26/2017 Cert. No.: 00025
Linearity File No.: 01748 Lin. Date: 12/26/2017 Lin. No.: 00025
Solution File No.: 01772 Soln. Date: 03/02/2018 Soln. No.: 00315
Sequential File No.: 01772 File Date: 03/02/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113
Control Solution %: 0.100% Expires: 12/11/2019
Solution Control Lot: 17420 Bottle No.: 1259

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:44S		
Control 1 EC	0.101%	13:45S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:45S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:46S		
Control 2 EC	0.100%	13:46S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	13:46S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:47S		
Control 3 EC	0.100%	13:48S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	13:48S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:48S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: 

Badge No.: 129

Date: 03/02/2018