

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01746 **Calib. Date:** 12/26/2017 **Calib. No.:** 00034
Certification File No.: 01747 **Cert. Date:** 12/26/2017 **Cert. No.:** 00025
Linearity File No.: 01748 **Lin. Date:** 12/26/2017 **Lin. No.:** 00025
Solution File No.: 01782 **Soln. Date:** 03/21/2018 **Soln. No.:** 00316
Sequential File No.: 01782 **File Date:** 03/21/2018

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXB S3-113
Control Solution %: 0.100% **Expires:** 12/11/2019
Solution Control Lot: 17420 **Bottle No.:** 1132

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:38S		
Control 1 EC	0.101%	10:39S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:39S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:39S		
Control 2 EC	0.100%	10:40S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	10:40S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:41S		
Control 3 EC	0.100%	10:41S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:41S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:42S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: 

Badge No.: 129

Date: 03/21/2018