

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01404 Calib. Date: 12/26/2017 Calib. No.: 00028
Certification File No.: 01405 Cert. Date: 12/26/2017 Cert. No.: 00024
Linearity File No.: 01406 Lin. Date: 12/26/2017 Lin. No.: 00024
Solution File No.: 01432 Soln. Date: 04/11/2018 Soln. No.: 00267
Sequential File No.: 01432 File Date: 04/11/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-114
Control Solution %: 0.100% Expires: 12/11/2019
Solution Control Lot: 17420 Bottle No.: 1198

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:30D		
Control 1 EC	0.101%	09:30D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:31D		
Control 2 EC	0.100%	09:32D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:33D		
Control 3 EC	0.099%	09:33D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

Badge No.: 129

Date: 04/11/2018