

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01404	Calib. Date:	12/26/2017
Certification File No.:	01405	Cert. Date:	12/26/2017
Linearity File No.:	01406	Lin. Date:	12/26/2017
Solution File No.:	01441	Soln. Date:	05/25/2018
Sequential File No.:	01441	File Date:	05/25/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-114
Solution Control Lot:	17420	Expires:	12/11/2019
		Bottle No.:	1195

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:20D		
Control 1 EC	0.101%	09:21D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:21D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:22D		
Control 2 EC	0.100%	09:22D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	09:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:23D		
Control 3 EC	0.100%	09:24D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: 

Badge No.: 129

Date: 05/25/2018