

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0037  
**Location:** WALL TOWNSHIP POLICE  
**Calibration File No.:** 01746 **Calib. Date:** 12/26/2017 **Calib. No.:** 00034  
**Certification File No.:** 01747 **Cert. Date:** 12/26/2017 **Cert. No.:** 00025  
**Linearity File No.:** 01748 **Lin. Date:** 12/26/2017 **Lin. No.:** 00025  
**Solution File No.:** 01808 **Soln. Date:** 06/13/2018 **Soln. No.:** 00320  
**Sequential File No.:** 01808 **File Date:** 06/13/2018

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDXB S3-113  
**Control Solution %:** 0.100% **Expires:** 12/11/2019  
**Solution Control Lot:** 17420 **Bottle No.:** 1122

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 09:34D        |                               |                           |
| Control 1 EC      | 0.101%         | 09:35D        | 33.9°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.100%         | 09:35D        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 09:36D        |                               |                           |
| Control 2 EC      | 0.100%         | 09:36D        | 33.9°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.101%         | 09:36D        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 09:37D        |                               |                           |
| Control 3 EC      | 0.099%         | 09:38D        | 33.9°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.102%         | 09:38D        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 09:38D        |                               |                           |

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: 

Badge No.: 129

Date: 06/13/2018