

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0037  
**Location:** WALL TOWNSHIP POLICE  
**Calibration File No.:** 01810 **Calib. Date:** 06/15/2018 **Calib. No.:** 00035  
**Certification File No.:** 01811 **Cert. Date:** 06/15/2018 **Cert. No.:** 00026  
**Linearity File No.:** 01812 **Lin. Date:** 06/15/2018 **Lin. No.:** 00026  
**Solution File No.:** 01817 **Soln. Date:** 06/28/2018 **Soln. No.:** 00322  
**Sequential File No.:** 01817 **File Date:** 06/28/2018

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDAB-0015  
**Control Solution %:** 0.100% **Expires:** 12/11/2019  
**Solution Control Lot:** 17420 **Bottle No.:** 1110

| Function          | Result | Time   | Temperature    | Comment(s)          |
|-------------------|--------|--------|----------------|---------------------|
|                   | %BAC   | HH:MM  | Simulator (°C) | or Error(s)         |
| Ambient Air Blank | 0.000% | 13:56D |                |                     |
| Control 1 EC      | 0.101% | 13:57D | 33.9°C         | *** TEST PASSED *** |
| Control 1 IR      | 0.100% | 13:57D | 33.9°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 13:58D |                |                     |
| Control 2 EC      | 0.100% | 13:58D | 33.9°C         | *** TEST PASSED *** |
| Control 2 IR      | 0.100% | 13:58D | 33.9°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 13:59D |                |                     |
| Control 3 EC      | 0.099% | 14:00D | 33.9°C         | *** TEST PASSED *** |
| Control 3 IR      | 0.100% | 14:00D | 33.9°C         | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 14:00D |                |                     |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A.

Signature: \_\_\_\_\_

PTL. K #129

Badge No.: 129

Date: 06/28/2018