

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01810 **Calib. Date:** 06/15/2018 **Calib. No.:** 00035
Certification File No.: 01811 **Cert. Date:** 06/15/2018 **Cert. No.:** 00026
Linearity File No.: 01812 **Lin. Date:** 06/15/2018 **Lin. No.:** 00026
Solution File No.: 01825 **Soln. Date:** 07/22/2018 **Soln. No.:** 00324
Sequential File No.: 01825 **File Date:** 07/22/2018

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0015
Control Solution %: 0.100% **Expires:** 12/11/2019
Solution Control Lot: 17420 **Bottle No.:** 1121

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:42D		
Control 1 EC	0.100%	20:42D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:42D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:43D		
Control 2 EC	0.099%	20:44D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	20:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:44D		
Control 3 EC	0.097%	20:45D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	20:45D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:45D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BALDINO

First Name: ANDREW

MI: M

Signature: _____

 #174

Badge No.: 174

Date: 07/22/2018

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: WALL TOWNSHIP POLICE
Serial No.: ARXA-0037
Calibration File No.: 01810 Calib. Date: 06/15/2018 Calib. No.: 00035
Certification File No.: 01811 Cert. Date: 06/15/2018 Cert. No.: 00026
Linearity File No.: 01812 Lin. Date: 06/15/2018 Lin. No.: 00026
Solution File No.: 01824 Soln. Date: 07/22/2018 Soln. No.: 00323
Sequential File No.: 01824 File Date: 07/22/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0015
Control Solution %: 0.100% Expires: 12/11/2019
Solution Control Lot: 17420 Bottle No.: 1190

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:38D		
Control 1 EC	-.---%	07:39D	33.9°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	07:39D	33.9°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BALDINO

First Name: ANDREW

MI: M

Signature: 

Badge No.: 174

Date: 07/22/2018