

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01453 Calib. Date: 06/15/2018 Calib. No.: 00029
Certification File No.: 01454 Cert. Date: 06/15/2018 Cert. No.: 00025
Linearity File No.: 01455 Lin. Date: 06/15/2018 Lin. No.: 00025
Solution File No.: 01474 Soln. Date: 07/22/2018 Soln. No.: 00273
Sequential File No.: 01474 File Date: 07/22/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0014
Control Solution %: 0.100% Expires: 12/11/2019
Solution Control Lot: 17420 Bottle No.: 1185

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:40D		
Control 1 EC	0.100%	07:41D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:41D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:42D		
Control 2 EC	0.099%	07:42D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:42D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:43D		
Control 3 EC	0.097%	07:44D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	07:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:44D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BALDINO

First Name: ANDREW

MI: M

Signature: 

Badge No.: 174

Date: 07/22/2018