

# Calibrating Unit

## New Standard Solution Report

**Equipment**

Alcotest 7110 MKIII-C  
Location: WALL TOWNSHIP POLICE  
Calibration File No.: 01810  
Certification File No.: 01811  
Linearity File No.: 01812  
Solution File No.: 01835  
Sequential File No.: 01835

Serial No.: ARXA-0037

Calib. Date: 06/15/2018  
Cert. Date: 06/15/2018  
Lin. Date: 06/15/2018  
Soln. Date: 08/26/2018  
File Date: 08/26/2018

Calib. No.: 00035  
Cert. No.: 00026  
Lin. No.: 00026  
Soln. No.: 00326

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 18150

Model No.: CU-34

Serial No.: DDAB-0015  
Expires: 05/10/2020  
Bottle No.: 0571

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 12:24D        |                               |                           |
| Control 1 EC      | 0.100%         | 12:24D        | 33.9°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.100%         | 12:24D        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 12:25D        |                               |                           |
| Control 2 EC      | 0.097%         | 12:26D        | 33.9°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.100%         | 12:26D        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 12:26D        |                               |                           |
| Control 3 EC      | 0.097%         | 12:27D        | 33.9°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.100%         | 12:27D        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 12:28D        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DRACKWICZ

First Name: AMY

MI: M

Signature: \_\_\_\_\_

Badge No.: 177

Date: 08/26/2018