

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01453 **Calib. Date:** 06/15/2018 **Calib. No.:** 00029
Certification File No.: 01454 **Cert. Date:** 06/15/2018 **Cert. No.:** 00025
Linearity File No.: 01455 **Lin. Date:** 06/15/2018 **Lin. No.:** 00025
Solution File No.: 01489 **Soln. Date:** 08/26/2018 **Soln. No.:** 00275
Sequential File No.: 01489 **File Date:** 08/26/2018

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0014
Control Solution %: 0.100% **Expires:** 12/11/2019
Solution Control Lot: 17420 **Bottle No.:** 1106

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:52D		
Control 1 EC	-.---%	11:53D	22.5°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	11:53D	22.5°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DRACKWICZ

First Name: AMY

MI: M

Signature: Agt [Signature] 177

Badge No.: 177

Date: 08/26/2018