

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01810 Calib. Date: 06/15/2018 Calib. No.: 00035
Certification File No.: 01811 Cert. Date: 06/15/2018 Cert. No.: 00026
Linearity File No.: 01812 Lin. Date: 06/15/2018 Lin. No.: 00026
Solution File No.: 01837 Soln. Date: 08/30/2018 Soln. No.: 00327
Sequential File No.: 01837 File Date: 08/30/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0015
Control Solution %: 0.100% Expires: 12/11/2019
Solution Control Lot: 17420 Bottle No.: 1222

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:18D		
Control 1 EC	0.100%	09:19D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:19D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:20D		
Control 2 EC	0.098%	09:20D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:20D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:21D		
Control 3 EC	0.098%	09:22D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:22D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:22D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: RLK #129

Badge No.: 129

Date: 08/30/2018