

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01453	Calib. Date:	06/15/2018
Certification File No.:	01454	Cert. No.:	00029
Linearity File No.:	01455	Lin. Date:	06/15/2018
Solution File No.:	01492	Cert. No.:	00025
Sequential File No.:	01492	Lin. No.:	00025
		Soln. Date:	08/30/2018
		Soln. No.:	00277
		File Date:	08/30/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	17420	Expires:	12/11/2019
		Bottle No.:	1179

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:25D		
Control 1 EC	0.101%	09:26D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:26D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:27D		
Control 2 EC	0.100%	09:27D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	09:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:28D		
Control 3 EC	0.099%	09:29D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRADY

First Name: PATRICK

MI: R

Signature: Pat. Brady #175

Badge No.: 175

Date: 08/30/2018