

Calibrating Unit New Standard Solution Report

| | | |
|-------------------------|-----------------------|-------------------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: ARXA-0037 |
| Location: | WALL TOWNSHIP POLICE | |
| Calibration File No.: | 01810 | Calib. Date: 06/15/2018 |
| Certification File No.: | 01811 | Calib. No.: 00035 |
| Linearity File No.: | 01812 | Cert. Date: 06/15/2018 |
| Solution File No.: | 01838 | Cert. No.: 00026 |
| Sequential File No.: | 01838 | Lin. Date: 06/15/2018 |
| | | Lin. No.: 00026 |
| | | Soln. Date: 09/14/2018 |
| | | Soln. No.: 00328 |
| | | File Date: 09/14/2018 |
| | | |
| Calibrating Unit: | WET | Model No.: CU-34 |
| Control Solution %: | 0.100% | Serial No.: DDAB-0015 |
| Solution Control Lot: | 18150 | Expires: 05/10/2020 |
| | | Bottle No.: 0507 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 12:27D | | |
| Control 1 EC | 0.100% | 12:27D | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 12:27D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:28D | | |
| Control 2 EC | 0.099% | 12:29D | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.099% | 12:29D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:29D | | |
| Control 3 EC | 0.097% | 12:30D | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 12:30D | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 12:31D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: _____

PTL. K #129

Badge No.: 129

Date: 09/14/2018