

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01453	Calib. Date:	06/15/2018
Certification File No.:	01454	Cert. No.:	00029
Linearity File No.:	01455	Cert. Date:	06/15/2018
Solution File No.:	01500	Lin. Date:	06/15/2018
Sequential File No.:	01500	Soln. Date:	09/14/2018
		Soln. No.:	00278
		File Date:	09/14/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	18150	Expires:	05/10/2020
		Bottle No.:	0339

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:31D		
Control 1 EC	0.097%	12:32D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	12:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:32D		
Control 2 EC	0.097%	12:33D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	12:33D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:34D		
Control 3 EC	0.096%	12:34D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	12:34D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:35D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: _____

PH. K #129

Badge No.: 129

Date: 09/14/2018