

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01453 Calib. Date: 06/15/2018 Calib. No.: 00029
Certification File No.: 01454 Cert. Date: 06/15/2018 Cert. No.: 00025
Linearity File No.: 01455 Lin. Date: 06/15/2018 Lin. No.: 00025
Solution File No.: 01511 Soln. Date: 09/28/2018 Soln. No.: 00281
Sequential File No.: 01511 File Date: 09/28/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0014
Control Solution %: 0.100% Expires: 05/10/2020
Solution Control Lot: 18150 Bottle No.: 0505

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:39D		
Control 1 EC	0.100%	12:39D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:40D		
Control 2 EC	0.099%	12:41D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:41D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:41D		
Control 3 EC	0.099%	12:42D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	12:42D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:43D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRADY

First Name: PATRICK

MI: R

Signature: Pat. Brady #175

Badge No.: 175

Date: 09/28/2018