

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01810 **Calib. Date:** 06/15/2018 **Calib. No.:** 00035
Certification File No.: 01811 **Cert. Date:** 06/15/2018 **Cert. No.:** 00026
Linearity File No.: 01812 **Lin. Date:** 06/15/2018 **Lin. No.:** 00026
Solution File No.: 01845 **Soln. Date:** 10/04/2018 **Soln. No.:** 00330
Sequential File No.: 01845 **File Date:** 10/04/2018

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0015
Control Solution %: 0.100% **Expires:** 05/10/2020
Solution Control Lot: 18150 **Bottle No.:** 0445

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:05D		
Control 1 EC	0.101%	15:05D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:05D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:06D		
Control 2 EC	0.098%	15:07D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:07D		
Control 3 EC	0.097%	15:08D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: -

Signature: PTL MCKAY #178

Badge No.: 178

Date: 10/04/2018