

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01810 **Calib. Date:** 06/15/2018 **Calib. No.:** 00035
Certification File No.: 01811 **Cert. Date:** 06/15/2018 **Cert. No.:** 00026
Linearity File No.: 01812 **Lin. Date:** 06/15/2018 **Lin. No.:** 00026
Solution File No.: 01849 **Soln. Date:** 10/19/2018 **Soln. No.:** 00331
Sequential File No.: 01849 **File Date:** 10/19/2018

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0015
Control Solution %: 0.100% **Expires:** 05/10/2020
Solution Control Lot: 18150 **Bottle No.:** 0619

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:22D		
Control 1 EC	0.099%	09:23D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:23D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:24D		
Control 2 EC	0.097%	09:24D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:25D		
Control 3 EC	0.097%	09:25D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:25D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:26D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: PTL #129

Badge No.: 129

Date: 10/19/2018