

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01453	Calib. Date:	06/15/2018
Certification File No.:	01454	Cert. Date:	06/15/2018
Linearity File No.:	01455	Lin. Date:	06/15/2018
Solution File No.:	01520	Soln. Date:	10/19/2018
Sequential File No.:	01520	File Date:	10/19/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	18150	Expires:	05/10/2020
		Bottle No.:	0277

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:26D		
Control 1 EC	0.096%	09:27D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	09:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:27D		
Control 2 EC	0.096%	09:28D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	09:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:29D		
Control 3 EC	0.096%	09:29D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:30D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: PTL. K. #129

Badge No.: 129

Date: 10/19/2018