

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01453	Calib. Date:	06/15/2018
Certification File No.:	01454	Cert. Date:	06/15/2018
Linearity File No.:	01455	Lin. Date:	06/15/2018
Solution File No.:	01523	Soln. Date:	10/31/2018
Sequential File No.:	01523	File Date:	10/31/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0014
Solution Control Lot:	18150	Expires:	05/10/2020
		Bottle No.:	0508

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:36S		
Control 1 EC	0.097%	07:37S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	07:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:37S		
Control 2 EC	0.096%	07:38S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	07:38S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:39S		
Control 3 EC	0.097%	07:39S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	07:39S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:40S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BALDINO

First Name: ANDREW

MI: M

Signature: _____



Badge No.: 174

Date: 10/31/2018