

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01858	Calib. Date:	11/16/2018
Certification File No.:	01859	Calib. No.:	00036
Linearity File No.:	01860	Cert. Date:	11/16/2018
Solution File No.:	01866	Lin. Date:	11/16/2018
Sequential File No.:	01866	Lin. No.:	00027
		Soln. Date:	12/26/2018
		Soln. No.:	00337
		File Date:	12/26/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-114
Solution Control Lot:	18290	Expires:	09/17/2020
		Bottle No.:	0130

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:09S		
Control 1 EC	0.099%	08:10S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	08:10S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:11S		
Control 2 EC	0.097%	08:11S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	08:11S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:12S		
Control 3 EC	0.096%	08:13S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:13S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:13S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PTL #129

Badge No.: 129

Date: 12/26/2018