

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01858 Calib. Date: 11/16/2018 Calib. No.: 00036
Certification File No.: 01859 Cert. Date: 11/16/2018 Cert. No.: 00027
Linearity File No.: 01860 Lin. Date: 11/16/2018 Lin. No.: 00027
Solution File No.: 01870 Soln. Date: 01/07/2019 Soln. No.: 00338
Sequential File No.: 01870 File Date: 01/07/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-114
Control Solution %: 0.100% Expires: 09/17/2020
Solution Control Lot: 18290 Bottle No.: 0764

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:07S		
Control 1 EC	0.098%	16:08S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	16:08S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:08S		
Control 2 EC	0.097%	16:09S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:10S		
Control 3 EC	0.095%	16:10S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:10S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:11S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: 

Badge No.: 129

Date: 01/07/2019