

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: WALL TOWNSHIP POLICE
Serial No.: ARXA-0037
Calibration File No.: 01858 Calib. Date: 11/16/2018 Calib. No.: 00036
Certification File No.: 01859 Cert. Date: 11/16/2018 Cert. No.: 00027
Linearity File No.: 01860 Lin. Date: 11/16/2018 Lin. No.: 00027
Solution File No.: 01873 Soln. Date: 01/10/2019 Soln. No.: 00339
Sequential File No.: 01873 File Date: 01/10/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-114
Control Solution %: 0.100% Expires: 09/17/2020
Solution Control Lot: 18290 Bottle No.: 0613

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:59S		
Control 1 EC	0.099%	16:00S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:00S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:01S		
Control 2 EC	0.097%	16:01S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:01S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:02S		
Control 3 EC	0.096%	16:03S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:03S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:03S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with
Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: I

Signature: _____

PTL. MCKAY #178

Badge No.: 178

Date: 01/10/2019