

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARZL-0146
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01528	Calib. Date:	11/16/2018
Certification File No.:	01529	Cert. Date:	11/16/2018
Linearity File No.:	01530	Lin. Date:	11/16/2018
Solution File No.:	01566	Soln. Date:	01/24/2019
Sequential File No.:	01566	File Date:	01/24/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-113
Solution Control Lot:	18290	Expires:	09/17/2020
		Bottle No.:	0593

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:14S		
Control 1 EC	0.100%	13:15S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	13:15S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:15S		
Control 2 EC	0.099%	13:16S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:16S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:17S		
Control 3 EC	0.100%	13:18S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	13:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL.

MI: A

Signature: _____

PTL. K #129

Badge No.: 129

Date: 01/24/2019