

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: WALL TOWNSHIP POLICE
Serial No.: ARXA-0037
Calibration File No.: 01858
Calib. Date: 11/16/2018
Calib. No.: 00036
Certification File No.: 01859
Cert. Date: 11/16/2018
Cert. No.: 00027
Linearity File No.: 01860
Lin. Date: 11/16/2018
Lin. No.: 00027
Solution File No.: 01879
Soln. Date: 01/30/2019
Soln. No.: 00342
Sequential File No.: 01879
File Date: 01/30/2019

Calibrating Unit: WET
Model No.: CU-34
Serial No.: DDXB S3-114
Control Solution %: 0.100%
Expires: 09/17/2020
Solution Control Lot: 18290
Bottle No.: 1285

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:35S		
Control 1 EC	0.102%	12:35S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:35S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:36S		
Control 2 EC	0.099%	12:37S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	12:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:38S		
Control 3 EC	0.098%	12:38S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:38S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:39S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: 

Badge No.: 129

Date: 01/30/2019