

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARZL-0146  
Location: WALL TOWNSHIP POLICE  
Calibration File No.: 01528 Calib. Date: 11/16/2018 Calib. No.: 00030  
Certification File No.: 01529 Cert. Date: 11/16/2018 Cert. No.: 00026  
Linearity File No.: 01530 Lin. Date: 11/16/2018 Lin. No.: 00026  
Solution File No.: 01570 Soln. Date: 01/30/2019 Soln. No.: 00294  
Sequential File No.: 01570 File Date: 01/30/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113  
Control Solution %: 0.100% Expires: 09/17/2020  
Solution Control Lot: 18290 Bottle No.: 0591

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 12:36S        |                               |                           |
| Control 1 EC      | 0.101%         | 12:36S        | 33.9°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.100%         | 12:36S        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 12:37S        |                               |                           |
| Control 2 EC      | 0.099%         | 12:38S        | 33.9°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.099%         | 12:38S        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 12:38S        |                               |                           |
| Control 3 EC      | 0.099%         | 12:39S        | 34.0°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.099%         | 12:39S        | 34.0°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 12:40S        |                               |                           |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: \_\_\_\_\_

*RLK #129*

Badge No.: 129

Date: 01/30/2019