

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01528 Calib. Date: 11/16/2018 Calib. No.: 00030
Certification File No.: 01529 Cert. Date: 11/16/2018 Cert. No.: 00026
Linearity File No.: 01530 Lin. Date: 11/16/2018 Lin. No.: 00026
Solution File No.: 01576 Soln. Date: 02/10/2019 Soln. No.: 00295
Sequential File No.: 01576 File Date: 02/10/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113
Control Solution %: 0.100% Expires: 09/17/2020
Solution Control Lot: 18290 Bottle No.: 0252

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:02S		
Control 1 EC	0.097%	23:02S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	23:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:03S		
Control 2 EC	0.096%	23:04S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	23:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:04S		
Control 3 EC	0.096%	23:05S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:06S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DRACKWICZ

First Name: AMY

MI: M

Signature: 

Badge No.: 177

Date: 02/10/2019