

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01858	Calib. Date:	11/16/2018
Certification File No.:	01859	Cert. Date:	11/16/2018
Linearity File No.:	01860	Lin. Date:	11/16/2018
Solution File No.:	01883	Soln. Date:	02/11/2019
Sequential File No.:	01883	File Date:	02/11/2019
		Calib. No.:	00036
		Cert. No.:	00027
		Lin. No.:	00027
		Soln. No.:	00344
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDXB S3-114
Solution Control Lot:	18290	Expires:	09/17/2020
		Bottle No.:	0248

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:55S		
Control 1 EC	0.099%	01:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	01:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:56S		
Control 2 EC	0.096%	01:57S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	01:57S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:58S		
Control 3 EC	0.095%	01:58S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	01:58S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:59S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DRACKWICZ

First Name: AMY

MI: M

Signature: _____

Badge No.: 177

Date: 02/11/2019