

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01858 Calib. Date: 11/16/2018 Calib. No.: 00036
Certification File No.: 01859 Cert. Date: 11/16/2018 Cert. No.: 00027
Linearity File No.: 01860 Lin. Date: 11/16/2018 Lin. No.: 00027
Solution File No.: 01884 Soln. Date: 02/21/2019 Soln. No.: 00345
Sequential File No.: 01884 File Date: 02/21/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-114
Control Solution %: 0.100% Expires: 01/08/2021
Solution Control Lot: 19020 Bottle No.: 1181

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:27S		
Control 1 EC	0.102%	10:27S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:28S		
Control 2 EC	0.099%	10:29S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:29S		
Control 3 EC	0.099%	10:30S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:31S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: I

Signature: PTL. MCKAY #178

Badge No.: 178

Date: 02/21/2019