

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C

Serial No.: ARXA-0037

Location:

WALL TOWNSHIP POLICE

Calibration File No.: 01858

Calib. Date: 11/16/2018

Calib. No.: 00036

Certification File No.: 01859

Cert. Date: 11/16/2018

Cert. No.: 00027

Linearity File No.: 01860

Lin. Date: 11/16/2018

Lin. No.: 00027

Solution File No.: 01888

Soln. Date: 02/24/2019

Soln. No.: 00346

Sequential File No.: 01888

File Date: 02/24/2019

Calibrating Unit: WET

Model No.: CU-34

Serial No.: DDXB S3-114

Control Solution %: 0.100%

Expires: 01/08/2021

Solution Control Lot: 19020

Bottle No.: 0964

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	03:03S		
Control 1 EC	0.100%	03:03S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:04S		
Control 2 EC	0.098%	03:05S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	03:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:05S		
Control 3 EC	0.096%	03:06S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:06S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:07S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRADY

First Name: PATRICK

MI: R

Signature:

Pat. A. Brady #175

Badge No.: 175

Date: 02/24/2019