

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01858 **Calib. Date:** 11/16/2018 **Calib. No.:** 00036
Certification File No.: 01859 **Cert. Date:** 11/16/2018 **Cert. No.:** 00027
Linearity File No.: 01860 **Lin. Date:** 11/16/2018 **Lin. No.:** 00027
Solution File No.: 01899 **Soln. Date:** 03/25/2019 **Soln. No.:** 00348
Sequential File No.: 01899 **File Date:** 03/25/2019

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDXB S3-114
Control Solution %: 0.100% **Expires:** 01/08/2021
Solution Control Lot: 19020 **Bottle No.:** 1288

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:20S		
Control 1 EC	0.101%	11:21S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:21S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:22S		
Control 2 EC	0.099%	11:22S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:22S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:23S		
Control 3 EC	0.098%	11:24S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: I

Signature: _____

PTL. MCKAY #178

Badge No.: 178

Date: 03/25/2019