

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01528 Calib. Date: 11/16/2018 Calib. No.: 00030
Certification File No.: 01529 Cert. Date: 11/16/2018 Cert. No.: 00026
Linearity File No.: 01530 Lin. Date: 11/16/2018 Lin. No.: 00026
Solution File No.: 01591 Soln. Date: 03/25/2019 Soln. No.: 00298
Sequential File No.: 01591 File Date: 03/25/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113
Control Solution %: 0.100% Expires: 09/17/2020
Solution Control Lot: 18290 Bottle No.: 0126

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:11S		
Control 1 EC	0.100%	11:12S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:12S		
Control 2 EC	0.099%	11:13S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:14S		
Control 3 EC	0.099%	11:14S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MCKAY JR

First Name: ORIN

MI: I

Signature: _____

PTL. MCKAY #178

Badge No.: 178

Date: 03/25/2019