

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01528 Calib. Date: 11/16/2018 Calib. No.: 00030
Certification File No.: 01529 Cert. Date: 11/16/2018 Cert. No.: 00026
Linearity File No.: 01530 Lin. Date: 11/16/2018 Lin. No.: 00026
Solution File No.: 01598 Soln. Date: 04/07/2019 Soln. No.: 00299
Sequential File No.: 01598 File Date: 04/07/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXB S3-113
Control Solution %: 0.100% Expires: 09/17/2020
Solution Control Lot: 18290 Bottle No.: 0179

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:46D		
Control 1 EC	0.098%	22:47D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:48D		
Control 2 EC	0.099%	22:48D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:49D		
Control 3 EC	0.099%	22:50D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	22:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:50D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: CACCIATORE

First Name: MICHAEL

MI: J

Signature: 

Badge No.: 187

Date: 04/07/2019