

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARZL-0146
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01601 **Calib. Date:** 04/17/2019 **Calib. No.:** 00031
Certification File No.: 01602 **Cert. Date:** 04/17/2019 **Cert. No.:** 00027
Linearity File No.: 01603 **Lin. Date:** 04/17/2019 **Lin. No.:** 00027
Solution File No.: 01613 **Soln. Date:** 05/04/2019 **Soln. No.:** 00301
Sequential File No.: 01613 **File Date:** 05/04/2019

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0014
Control Solution %: 0.100% **Expires:** 09/17/2020
Solution Control Lot: 18290 **Bottle No.:** 0209

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:41D		
Control 1 EC	0.098%	06:41D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	06:41D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:42D		
Control 2 EC	0.098%	06:43D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	06:43D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:43D		
Control 3 EC	0.097%	06:44D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:45D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: JENNINGS

First Name: ERIC

MI: R

Signature: 

Badge No.: 168

Date: 05/04/2019