

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARXA-0037  
Location: WALL TOWNSHIP POLICE  
Calibration File No.: 01905 Calib. Date: 04/17/2019 Calib. No.: 00037  
Certification File No.: 01906 Cert. Date: 04/17/2019 Cert. No.: 00028  
Linearity File No.: 01907 Lin. Date: 04/17/2019 Lin. No.: 00028  
Solution File No.: 01922 Soln. Date: 05/17/2019 Soln. No.: 00352  
Sequential File No.: 01922 File Date: 05/17/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0015  
Control Solution %: 0.100% Expires: 01/08/2021  
Solution Control Lot: 19020 Bottle No.: 0963

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:10D		
Control 1 EC	0.099%	05:10D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	05:10D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:11D		
Control 2 EC	0.096%	05:12D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	05:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:13D		
Control 3 EC	0.096%	05:13D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	05:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:14D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: BALDINO

First Name: ANDREW

MI: M

Signature: \_\_\_\_\_



Badge No.: 174

Date: 05/17/2019