

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01905 **Calib. Date:** 04/17/2019 **Calib. No.:** 00037
Certification File No.: 01906 **Cert. Date:** 04/17/2019 **Cert. No.:** 00028
Linearity File No.: 01907 **Lin. Date:** 04/17/2019 **Lin. No.:** 00028
Solution File No.: 01927 **Soln. Date:** 05/25/2019 **Soln. No.:** 00354
Sequential File No.: 01927 **File Date:** 05/25/2019

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDAB-0015
Control Solution %: 0.100% **Expires:** 01/08/2021
Solution Control Lot: 19020 **Bottle No.:** 0910

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:31D		
Control 1 EC	0.095%	05:32D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	05:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:33D		
Control 2 EC	0.094%	05:34D	33.9°C	CONTROL TEST FAILED
Control 2 IR	0.098%	05:34D	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	05:34D		
Control 3 EC	0.095%	05:35D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	05:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:36D		

Test results are not within acceptable range.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: SANTORO

First Name: DANIEL

MI: J

Signature: 

Badge No.: 165

Date: 05/25/2019