

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARXA-0037
Location:	WALL TOWNSHIP POLICE		
Calibration File No.:	01905	Calib. Date:	04/17/2019
Certification File No.:	01906	Cert. Date:	04/17/2019
Linearity File No.:	01907	Lin. Date:	04/17/2019
Solution File No.:	01929	Soln. Date:	05/26/2019
Sequential File No.:	01929	File Date:	05/26/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAB-0015
Solution Control Lot:	19020	Expires:	01/08/2021
		Bottle No.:	1017

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:39D		
Control 1 EC	0.099%	09:40D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	09:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:41D		
Control 2 EC	0.097%	09:41D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	09:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:42D		
Control 3 EC	0.096%	09:43D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:44D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BALDINO

First Name: ANDREW

MI: M

Signature: _____



Badge No.: 174

Date: 05/26/2019