

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARXA-0037
Location: WALL TOWNSHIP POLICE
Calibration File No.: 01905 Calib. Date: 04/17/2019 Calib. No.: 00037
Certification File No.: 01906 Cert. Date: 04/17/2019 Cert. No.: 00028
Linearity File No.: 01907 Lin. Date: 04/17/2019 Lin. No.: 00028
Solution File No.: 01934 Soln. Date: 05/28/2019 Soln. No.: 00359
Sequential File No.: 01934 File Date: 05/28/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAB-0015
Control Solution %: 0.100% Expires: 01/08/2021
Solution Control Lot: 19020 Bottle No.: 1393

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:43D		
Control 1 EC	0.098%	13:43D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	13:43D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:44D		
Control 2 EC	0.096%	13:45D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	13:45D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:46D		
Control 3 EC	0.095%	13:46D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	13:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KAZANOWSKY

First Name: PTL

MI: A

Signature: _____

PLK H129

Badge No.: 129

Date: 05/28/2019